

CERTIFICATE OF ANALYSIS

PRODUCT CM0361B
M.R.S. AGAR 500g

LOT NUMBER 6275633

EXPIRY DATE 2027.06.30

**DATE OF
MANUFACTURE** 2025.06.02

Delivery/Customer information

Date Printed

2025.06.25

Delivery No.

Customer

Customer Order Number

Physical Characteristics

Results

Specification

Appearance
Colour on reconstitution
pH (25°C)
Clarity

Dark straw powder
Straw 5-Amber 1
6.0
Clear

Dark straw powder
Straw 5-Amber 1
6.0 - 6.4
Clear

Microbiological Performance

Control
cfu

Test
cfu

Recovery Test
% Description

Microaerophilic incubation at 30°C for 72 hours

<i>Lactobacillus sakei</i> ATCC®15521	65	59	91	Pale straw cols
<i>Lactobacillus gasseri</i> ATCC®19992	30	36	120	Pale straw cols
<i>Lactococcus lactis</i> ATCC®19435	48	46	96	Pale straw cols
<i>Pediococcus pentosaceus</i> ATCC®33316	13	13	100	Pale straw cols

Control Medium: MRS Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu).
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.23

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.