

CERTIFICATE OF ANALYSIS

PRODUCT CM0485B
VIOLET RED BILE GLUCOSE AGAR 500g

LOT NUMBER 6262749

EXPIRY DATE 2030.05.31

DATE OF MANUFACTURE 2025.04.25

Delivery/Customer information

Date Printed

2025.05.17

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw/pink powder	Straw/pink powder
Colour on reconstitution	Purple	Purple
pH (25°C)	7.5	7.2 - 7.6
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
Aerobic incubation at 37 ± 2°C for 24 ± 2 hours				
Pour plate technique				
<i>Klebsiella pneumoniae</i> ATCC®29665	108	100	93	Purple/pink cols & halo
<i>Proteus mirabilis</i> ATCC®12453	105	84	80	Purple/pink cols & halo
Surface plate technique				
<i>Shigella sonnei</i> ATCC®25931	87	73	84	Irregular, purple/pink cols & halo
<i>Enterobacter aerogenes</i> ATCC®13048	98	85	87	Purple/pink, mucoid cols & halo
<i>Pseudomonas aeruginosa</i> ATCC®9027	58	47	82	Straw cols, no halo
<i>Staphylococcus aureus</i> ATCC®6538	>1E+04	0	-	No growth

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at 37 ± 2°C for 24 ± 2 hours

Pour plate technique

<i>E.coli</i> ATCC®25922 WDCM00013	96	79	82	Purple/pink cols & halo
<i>E.coli</i> ATCC®8739 WDCM00012	75	73	97	Purple/pink cols & halo
<i>S.typhimurium</i> ATCC®14028 WDCM00031	56	61	109	Purple/pink cols & halo
<i>S.enteritidis</i> ATCC®13076 WDCM00030	73	69	95	Purple/pink cols & halo

Surface plate technique

<i>E.faecalis</i> ATCC®29212 WDCM00087	>1E+04	0	-	No growth
<i>E.faecalis</i> ATCC®19433 WDCM00009	>1E+04	0	-	No growth

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 50% of the control medium.

Refer to product specification for full details.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT	CM0485B
	VIOLET RED BILE GLUCOSE AGAR 500g
LOT NUMBER	6262749

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.12

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.