



CERTIFICATE OF ANALYSIS

PRODUCT CM0331B

COLUMBIA BLOOD AGAR BASE 500g

LOT NUMBER

6267416

EXPIRY DATE

2030.05.31

DATE OF

MANUFACTURE 2025.05.09

Delivery/Customer information

Date Printed 2025.06.03

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results		Specification		
Appearance	Straw power	Straw powder Straw 2		Straw powder	
Colour on reconstitution				Straw 2-3	
pH (25°C)	7.4 Clear Satisfactory		7.1 - 7.5 Clear Satisfactory		
Clarity					
Blood compatibility with horse blood					
Blood compatibility with sheep blood	Satisfactory		Satisfactory		
Microbiological Performance	Control cfu	Test cfu	Recovery	Test Description	
Aerobic incubation at 37°C for 24 hours					
Staphylococcus aureus ATCC®25923	63	64	102	Cream cols	
Anaerobic incubation at 37°C for 48 hours					
Clostridium sporogenes ATCC®19404	70	67	96	Pale straw cols	
Clostridium sporogenes ATCC®11437	68	68	100	Pale straw cols	
Enriched with 5% v/v horse blood					
Aerobic incubation at 37°C for 24 hours					
Streptococcus pyogenes ATCC®19615	41	39	95	Pale straw cols, beta haemolysis	
Streptococcus pneumoniae ATCC®6305	30	26	87	Grey/green cols, alpha haemolysis	
Enriched with 5% v/v horse blood					
Microaerophilic incubation at 37°C for 48 hours					
Neisseria gonorrhoeae NCTC11148	27	28	104	Grey/brown cols	
Growth of <i>Haemophilus</i> spp. in the presence of X Aerobic incubation at 37°C for 18 hours	& V factor dis	cs			
	X Factor (mm)	V Factor (mm)	X+V Factor (mm)		
Haemophilus influenzae ATCC®49247	0	0	18		
Haemophilus parainfluenzae ATCC®33392	0	45	49		





CERTIFICATE OF ANALYSIS

PRODUCT CM0331B

COLUMBIA BLOOD AGAR BASE 500g

LOT NUMBER 6267416

Enriched with 7% v/v horse blood

Bacitracin activity

Aerobic incubation at 37°C for 18 hours

	Zone Size	Limits	
	(mm)	(mm)	
Streptococcus pyogenes ATCC®19615	15	10 - 20	

Testing has been performed in accordance with ISO	11133:2014							
Enriched with 5% v/v sheep blood								
Microaerophilic incubation at $41.5 \pm 2^{\circ}$ C for 44 ± 4 hours								
C. jejuni ATCC®29428 WDCM00156	64	61	95	Grey, mucoid cols				
C. jejuni ATCC®33291 WDCM00005	109	95	87	Grey, mucoid cols				
C. coli ATCC®43478 WDCM00004	58	66	114	Grey, mucoid cols				
	Initial		Test Result					
	Inoculum							
Streptococcus pyogenes ATCC®19615	3400		Pale strav	v cols, beta haemolysis				
Streptococcus pneumoniae ATCC®6305	6000		Grey/green cols, alpha haemolysis					
Staphylococcus aureus ATCC®25923	1160		Cream cols					
Escherichia coli ATCC®25922	1140		Cream co	ls				

Control Medium: Columbia Blood Agar Base

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium. Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.02

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.