

## CERTIFICATE OF ANALYSIS

**PRODUCT** CM0331B  
COLUMBIA BLOOD AGAR BASE 500g

**LOT NUMBER** 6251256

**EXPIRY DATE** 2030.04.30

**DATE OF MANUFACTURE** 2025.04.02

## Delivery/Customer information

Date Printed

2025.05.12

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2-3	Straw 2-3
pH (25°C)	7.3	7.1 - 7.5
Clarity	Clear	Clear
Blood compatibility with horse blood	Satisfactory	Satisfactory
Blood compatibility with sheep blood	Satisfactory	Satisfactory

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
<b>Aerobic incubation at 37°C for 24 hours</b>				
<i>Staphylococcus aureus</i> ATCC®25923	72	64	89	Cream cols
<b>Anaerobic incubation at 37°C for 48 hours</b>				
<i>Clostridium sporogenes</i> ATCC®19404	84	85	101	Pale straw cols
<i>Clostridium sporogenes</i> ATCC®11437	81	85	105	Pale straw cols
<b>Enriched with 5% v/v horse blood</b>				
<b>Aerobic incubation at 37°C for 24 hours</b>				
<i>Streptococcus pyogenes</i> ATCC®19615	28	25	89	Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	29	32	110	Grey/green cols, alpha haemolysis
<b>Enriched with 5% v/v horse blood</b>				
<b>Microaerophilic incubation at 37°C for 48 hours</b>				
<i>Neisseria gonorrhoeae</i> NCTC11148	59	59	100	Grey/brown cols
<b>Growth of <i>Haemophilus</i> spp. in the presence of X &amp; V factor discs</b>				
<b>Aerobic incubation at 37°C for 18 hours</b>				
	<b>X Factor (mm)</b>	<b>V Factor (mm)</b>	<b>X+V Factor (mm)</b>	
<i>Haemophilus influenzae</i> ATCC®49247	0	0	20	
<i>Haemophilus parainfluenzae</i> ATCC®33392	0	49	49	

## CERTIFICATE OF ANALYSIS

**PRODUCT** CM0331B  
**COLUMBIA BLOOD AGAR BASE 500g**  
**LOT NUMBER** 6251256

**Enriched with 7% v/v horse blood**  
**Bacitracin activity**  
**Aerobic incubation at 37°C for 18 hours**

	<b>Zone Size (mm)</b>	<b>Limits (mm)</b>
<i>Streptococcus pyogenes</i> ATCC®19615	18	10 - 20

**Testing has been performed in accordance with ISO11133:2014**

**Enriched with 5% v/v sheep blood**

**Microaerophilic incubation at 41.5 ± 2°C for 44 ± 4 hours**

<i>C. jejuni</i> ATCC®29428 WDCM00156	59	63	107	Grey, mucoid cols
<i>C. jejuni</i> ATCC®33291 WDCM00005	64	56	88	Grey, mucoid cols
<i>C. coli</i> ATCC®43478 WDCM00004	69	66	96	Grey, mucoid cols
	<b>Initial Inoculum</b>		<b>Test Result</b>	
<i>Streptococcus pyogenes</i> ATCC®19615	2600			Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	5800			Grey/green cols, alpha haemolysis
<i>Staphylococcus aureus</i> ATCC®25923	1440			Cream cols
<i>Escherichia coli</i> ATCC®25922	780			Cream cols

Control Medium: Columbia Blood Agar Base

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.  
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.12

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.