

CERTIFICATE OF ANALYSIS

PRODUCT CM0331B
COLUMBIA BLOOD AGAR BASE 500g

LOT NUMBER 6271348

EXPIRY DATE 2030.05.31

DATE OF MANUFACTURE 2025.05.15

Delivery/Customer information

Date Printed

2025.06.13

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 2-3
pH (25°C)	7.3	7.1 - 7.5
Clarity	Clear	Clear
Blood compatibility with horse blood	Satisfactory	Satisfactory
Blood compatibility with sheep blood	Satisfactory	Satisfactory

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
Aerobic incubation at 37°C for 24 hours				
<i>Staphylococcus aureus</i> ATCC®25923	64	48	75	Cream cols
Anaerobic incubation at 37°C for 48 hours				
<i>Clostridium sporogenes</i> ATCC®19404	61	64	105	Pale straw cols
<i>Clostridium sporogenes</i> ATCC®11437	26	32	123	Pale straw cols
Enriched with 5% v/v horse blood				
Aerobic incubation at 37°C for 24 hours				
<i>Streptococcus pyogenes</i> ATCC®19615	31	28	90	Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	51	53	104	Grey/green cols, alpha haemolysis
Enriched with 5% v/v horse blood				
Microaerophilic incubation at 37°C for 48 hours				
<i>Neisseria gonorrhoeae</i> NCTC11148	32	32	100	Grey/brown cols
Growth of <i>Haemophilus</i> spp. in the presence of X & V factor discs				
Aerobic incubation at 37°C for 18 hours				
	X Factor (mm)	V Factor (mm)	X+V Factor (mm)	
<i>Haemophilus influenzae</i> ATCC®49247	0	0	21	
<i>Haemophilus parainfluenzae</i> ATCC®33392	0	51	47	

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT CM0331B
COLUMBIA BLOOD AGAR BASE 500g
LOT NUMBER 6271348

Enriched with 7% v/v horse blood
Bacitracin activity
Aerobic incubation at 37°C for 18 hours

	Zone Size (mm)	Limits (mm)
<i>Streptococcus pyogenes</i> ATCC®19615	17	10 - 20

Testing has been performed in accordance with ISO11133:2014

Enriched with 5% v/v sheep blood

Microaerophilic incubation at 41.5 ± 2°C for 44 ± 4 hours

<i>C. jejuni</i> ATCC®29428 WDCM00156	98	87	88	Grey, mucoid cols
<i>C. jejuni</i> ATCC®33291 WDCM00005	76	106	139	Grey, mucoid cols
<i>C. coli</i> ATCC®43478 WDCM00004	53	38	71	Grey, mucoid cols
	Initial Inoculum	Test Result		
<i>Streptococcus pyogenes</i> ATCC®19615	620	Pale straw cols, beta haemolysis		
<i>Streptococcus pneumoniae</i> ATCC®6305	1020	Grey/green cols, alpha haemolysis		
<i>Staphylococcus aureus</i> ATCC®25923	1280	Cream cols		
<i>Escherichia coli</i> ATCC®25922	920	Cream cols		

Control Medium: Columbia Blood Agar Base

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.13

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.