



CERTIFICATE OF ANALYSIS

PRODUCT CM0331B

COLUMBIA BLOOD AGAR BASE 500g

LOT NUMBER

6290624

EXPIRY DATE

2030.07.31

DATE OF

MANUFACTURE 2025.07.07

Delivery/Customer information

Date Printed 2025.07.23

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Results Straw powder		Specification		
Appearance	Straw power			Straw powder		
Colour on reconstitution	Straw 2		Straw 2-3			
pH (25°C)	7.3	7.3		7.1 - 7.5		
Clarity	Clear Satisfactory		Clear Satisfactory			
Blood compatibility with horse blood						
Blood compatibility with sheep blood	Satisfactory		Satisfactory			
Microbiological Performance	Control cfu	Test cfu	Recovery	Test Description		
		Ciu	70	Description		
Aerobic incubation at 37°C for 24 hours						
Staphylococcus aureus ATCC®25923	79	92	116	Cream cols		
Anaerobic incubation at 37°C for 48 hours						
Clostridium sporogenes ATCC®19404	91	69	76	Pale straw cols		
Clostridium sporogenes ATCC®11437	48	36	76	Pale straw cols		
Enriched with 5% v/v horse blood						
Aerobic incubation at 37°C for 24 hours						
Streptococcus pyogenes ATCC®19615	44	36	82	Pale straw cols, beta haemolysis		
Streptococcus pneumoniae ATCC®6305	43	44	102	Grey/green cols, alpha haemolysis		
Enriched with 5% v/v horse blood						
Microaerophilic incubation at 37°C for 48 hours						
Neisseria gonorrhoeae NCTC11148	41	43	105	Grey/brown cols		
Growth of <i>Haemophilus</i> spp. in the presence of X Aerobic incubation at 37°C for 18 hours	& V factor dis	scs				
	X Factor (mm)	V Factor (mm)	X+V Factor (mm)			
Haemophilus influenzae ATCC®49247	0	0	17			
Haemophilus parainfluenzae ATCC®33392	0	39	42			





CERTIFICATE OF ANALYSIS

PRODUCT CM0331B

COLUMBIA BLOOD AGAR BASE 500g

LOT NUMBER 6290624

Enriched with 7% v/v horse blood

Bacitracin activity

Aerobic incubation at 37°C for 18 hours

	Zone Size	Limits	
	(mm)	(mm)	
Streptococcus pyogenes ATCC®19615	18	10 - 20	

Testing has been performed in accordance with ISC	011133:2014							
Enriched with 5% v/v sheep blood								
Microaerophilic incubation at $41.5 \pm 2^{\circ}$ C for 44 ± 4 hours								
C. jejuni ATCC®29428 WDCM00156	109	95	87	Grey, mucoid cols				
C. jejuni ATCC®33291 WDCM00005	64	62	97	Grey, mucoid cols				
C. coli ATCC®43478 WDCM00004	62	60	97	Grey, mucoid cols				
	Initial		Test Result					
	Inoculum							
Streptococcus pyogenes ATCC®19615	8800		Pale stra	w cols, beta haemolysis				
Streptococcus pneumoniae ATCC®6305	8600		Grey/green cols, alpha haemolysis					
Staphylococcus aureus ATCC®25923	1520		Cream cols					
Escherichia coli ATCC®25922	1060		Cream c	ols				

Control Medium: Columbia Blood Agar Base

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium. Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.07.23

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.