

CERTIFICATE OF ANALYSIS

PRODUCT CM0331B
COLUMBIA BLOOD AGAR BASE 500g

LOT NUMBER 6293899

EXPIRY DATE 2030.07.31

DATE OF MANUFACTURE 2025.07.16

Delivery/Customer information

Date Printed

2025.08.05

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2-3	Straw 2-3
pH (25°C)	7.3	7.1 - 7.5
Clarity	Clear	Clear
Blood compatibility with horse blood	Satisfactory	Satisfactory
Blood compatibility with sheep blood	Satisfactory	Satisfactory

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
Aerobic incubation at 37°C for 24 hours				
<i>Staphylococcus aureus</i> ATCC®25923	46	55	119	Cream cols
Anaerobic incubation at 37°C for 48 hours				
<i>Clostridium sporogenes</i> ATCC®19404	54	56	104	Pale straw cols
<i>Clostridium sporogenes</i> ATCC®11437	36	37	103	Pale straw cols
Enriched with 5% v/v horse blood				
Aerobic incubation at 37°C for 24 hours				
<i>Streptococcus pyogenes</i> ATCC®19615	40	45	113	Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	11	12	109	Grey/green cols, alpha haemolysis
Enriched with 5% v/v horse blood				
Microaerophilic incubation at 37°C for 48 hours				
<i>Neisseria gonorrhoeae</i> NCTC11148	25	31	124	Grey/brown cols
Growth of <i>Haemophilus</i> spp. in the presence of X & V factor discs				
Aerobic incubation at 37°C for 18 hours				
	X Factor (mm)	V Factor (mm)	X+V Factor (mm)	
<i>Haemophilus influenzae</i> ATCC®49247	0	0	18	
<i>Haemophilus parainfluenzae</i> ATCC®33392	0	37	35	

CERTIFICATE OF ANALYSIS

PRODUCT CM0331B
COLUMBIA BLOOD AGAR BASE 500g
LOT NUMBER 6293899

Enriched with 7% v/v horse blood
Bacitracin activity
Aerobic incubation at 37°C for 18 hours

	Zone Size (mm)	Limits (mm)
<i>Streptococcus pyogenes</i> ATCC®19615	19	10 - 20

Testing has been performed in accordance with ISO11133:2014

Enriched with 5% v/v sheep blood

Microaerophilic incubation at 41.5 ± 2°C for 44 ± 4 hours

<i>C. jejuni</i> ATCC®29428 WDCM00156	65	69	106	Grey, mucoid cols
<i>C. jejuni</i> ATCC®33291 WDCM00005	72	78	108	Grey, mucoid cols
<i>C. coli</i> ATCC®43478 WDCM00004	84	81	96	Grey, mucoid cols
	Initial		Test Result	
	Inoculum			
<i>Streptococcus pyogenes</i> ATCC®19615	3600			Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	2200			Grey/green cols, alpha haemolysis
<i>Staphylococcus aureus</i> ATCC®25923	640			Cream cols
<i>Escherichia coli</i> ATCC®25922	1560			Cream cols

Control Medium: Columbia Blood Agar Base

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.08.04

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.