

CERTIFICATE OF ANALYSIS

PRODUCT CT0417B
CRO30 CEFTRIAXONE

LOT NUMBER 6323189

EXPIRY DATE 2028.09.24

DATE OF MANUFACTURE 2025.09.25

Delivery/Customer information

Date Printed

2025.10.02

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 29.1

Tested in accordance with CLSI current methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	26	22 - 28
<i>Escherichia coli</i> ATCC®25922	30	29 - 35
<i>Pseudomonas aeruginosa</i> ATCC®27853	21	17 - 23
<i>Haemophilus influenzae</i> ATCC®49247	33	31 - 39
<i>Streptococcus pneumoniae</i> ATCC®49619	35	30 - 35
<i>Neisseria gonorrhoeae</i> ATCC®49226	44	39 - 51

Tested in accordance with current EUCAST methodology

<i>Escherichia coli</i> ATCC®25922	30	29 - 35
<i>Klebsiella pneumoniae</i> ATCC®700603	18	16 - 22
<i>Streptococcus pneumoniae</i> ATCC®49619	34	32 - 38
<i>Haemophilus influenzae</i> ATCC®49766	38	34 - 42

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.10.02

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com