

Certificate of Analysis

Product Name: STREPTOCOCCAL GROUP REAGENT C 1X50 TESTS
Product Code: DR0588G
Lot Number: 6265662
Expiration Date: 03.06.2026
Date of Manufacture: 14.08.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR588A	Latex C	6235873	03.06.2026

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC	NCTC	Performance Criteria
S. pyogenes	936			Acceptable
S. agalactiae	761	9993		Acceptable
S. equisimilis	308	35666	8543	Acceptable
S. equisimilis	950	5371		Acceptable
S. equisimilis	951			Acceptable
S. equisimilis	953			Acceptable
S. faecalis	640	29212		Acceptable
S. faecalis	965			Acceptable
S. milleri	972			Acceptable
S. canis	975			Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
 Quality Assurance Manager