

Certificate of Analysis

Product Name: ESCHERICHIA COLI O157 LATEX TST 1X100TST
Product Code: DR0620M
Lot Number: 6205110
Expiration Date: 24.03.2026
Date of Manufacture: 30.04.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR621M	E.Coli Test Latex	6191275	24.03.2026
W74DR622M	E.Coli Control Latex Rgt	6191276	24.03.2026
W74DR623M	E.Coli Positive Control	6191277	25.03.2026
W74DR624M	E.Coli Negative Control	6191278	25.03.2026

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC	NCTC	Performance Criteria
E. coli O157	9155			Acceptable
E. coli O157	9153			Acceptable
E. coli O157	900			Acceptable
E. coli O157	901			Acceptable
E. coli O157	9161			Acceptable
E. coli O157	1872	700728	12900	Acceptable
E. coli O116	1818		9116	Acceptable
E. coli O130	130			Acceptable
E. coli	199	25922		Acceptable
E. coli	402	11775	9001	Acceptable
E. hermanii	892			Acceptable
E. hermanii	893			Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
Quality Assurance Manager