

## Certificate of Analysis

**Product Name:** PathoDextra Strep Grouping Kit  
**Product Code:** DR0700M  
**Lot Number:** 6238263  
**Expiration Date:** 31.03.2026  
**Date of Manufacture:** 29.05.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74PDX70A	Strep A Group Latex	6135357	31.05.2026
W74PDX71B	Strep B Group Latex	6194092	30.11.2026
W74PDX72C	Strep C Group Latex	6194091	30.11.2026
W74PDX73D	Strep D Group Latex	6238890	30.06.2026
W74PDX74F	Strep F Group Latex	6147419	30.04.2026
W74PDX75G	Strep G Group Latex	3808946	31.03.2026
W74PDX76H	Pathodextra Positive Control	6194087	28.02.2027
W74PDX77J	Reagent 1	6139163	30.06.2026
W74PDX78K	Reagent 2	6139164	30.06.2026
W74PDX79L	Reagent 3	6139165	30.06.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Latex Efficacy/ Specificity using Extraction Reagents	The Latex must show obvious agglutination with its homologous extract	Conforms
	The Latex must not show agglutination with the heterologous extracts but a trace reaction is acceptable	Conforms
Reactivity of Latex with Positive Control	The Latex must show strong positive agglutination with the positive control.	Conforms



Fay Curd  
Quality Assurance Manager