

CERTIFICATE OF ANALYSIS

PRODUCT SR0088E
PERFRINGENS (TSC) SELECT SUPP.1 X 10VIAL

LOT NUMBER 6249213

EXPIRY DATE 2027.04.14

DATE OF MANUFACTURE 2025.04.14

Delivery/Customer information

Date Printed

2025.05.06

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	White pellet	White pellet

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	-------------	----------	-----------------	-------------

Tested in CM587 with the addition of 5% v/v Egg Yolk Emulsion SR0047.

Anaerobic incubation at 37 ± 2°C for 16-24 hours

Pour plate technique

<i>Clostridium perfringens</i> ATCC®13124	34	36	106	Black cols with halo
<i>Clostridium bifermentans</i> NCTC506	99	89	90	Black cols with halo

Surface plate technique

<i>Clostridium tetani</i> ATCC®9441	>1E+04	0	-	No growth
<i>Proteus mirabilis</i> ATCC®29906	>1E+04	0	-	Translucent cols

Testing has been performed in accordance with ISO11133:2014

Table E

ISO Standard 7937 tested in Perfringens Agar Base (TSC/SFP) CM0587

Anaerobic incubation at 37 ± 2°C for 20 ± 2 hours

Pour plate technique

<i>C. perfringens</i> ATCC®13124 WDCM00007	52	50	96	Black cols
<i>C. perfringens</i> ATCC®12916 WDCM00080	71	77	108	Black cols

Surface plate technique

<i>E. coli</i> ATCC®25922 WDCM00013	>1E+04	0	-	No growth
<i>E. coli</i> ATCC®8739 WDCM00012	>1E+04	0	-	No growth

Table F

ISO Standard 14189 tested in Perfringens Agar Base (TSC/SFP) CM0587

Anaerobic incubation at 44 ± 2°C for 21 ± 3 hours

Filtration technique with overlay

<i>C. perfringens</i> ATCC®13124 WDCM00007	75	81	108	Black cols
<i>C. perfringens</i> ATCC®12916 WDCM00080	86	85	99	Black cols
<i>C. perfringens</i> ATCC®10543 WDCM00174	87	80	92	Black cols

Filtration technique without overlay

<i>B. subtilis</i> ATCC®6633 WDCM00003	2.5E+04	-	-	No growth
--	---------	---	---	-----------

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT	SR0088E
LOT NUMBER	PERFRINGENS (TSC) SELECT SUPP.1 X 10VIAL
	6249213

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by recovery of positive strains equal to or greater than 50% of the control medium.
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.06

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.