

CERTIFICATE OF ANALYSIS

PRODUCT SR0113E
BROAD SPECTRUM BETA-LACTAMASE MIX 1X10VL

LOT NUMBER 6259540

EXPIRY DATE 2026.12.13

DATE OF MANUFACTURE 2025.05.01

Delivery/Customer information

Date Printed

2025.05.07

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Potency from certificate β -lactamase I	1010	>500
Potency from certificate β -lactamase II	72	>50
Sterility test from certificate	Satisfactory	Satisfactory

Refer to product specification for full details.

A satisfactory sterility test result is represented by no detectable growth after incubation for 14 days at 20-25°C in Tryptone Soya Broth and at 30-35°C in Fluid Thioglycollate Medium.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.07

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.