

CERTIFICATE OF ANALYSIS

PRODUCT SR0186E
VANCOMYCIN SUPPLEMENT

LOT NUMBER 6263404

EXPIRY DATE 2027.04.22

**DATE OF
MANUFACTURE** 2025.04.22

Delivery/Customer information

Date Printed

2025.05.28

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	White pellet	White pellet

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

Tested in VRE Agar Base CM0985 with the addition of Meropenem Supplement SR0184

Aerobic incubation at 37°C for 48 hours

<i>Enterococcus faecalis</i> VRE NCTC12201	18	22	122	Black cols, aesculin hydrolysis
<i>Enterococcus faecium</i> NCTC12202	28	21	75	Black cols, aesculin hydrolysis
<i>Enterococcus faecalis</i> ATCC®33186	>1E+04	-	-	No growth
<i>Enterococcus faecium</i> ATCC®19434	>1E+04	-	-	No growth
<i>Escherichia coli</i> ATCC®25922	>1E+04	-	-	No growth
<i>Proteus mirabilis</i> NCTC10975	>1E+04	-	-	No growth
<i>Leuconostoc mesenteroides</i> ATCC®10830a	>1E+04	-	-	Black cols, aesculin hydrolysis

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). For positive test strains inoculated using diminishing sweep technique, a satisfactory result is represented by growth in accordance with the specification from an inoculum of 1E+04 to 1E+06 cfu. Negative strains are inhibited from an inoculum of 1E+04 to 1E+06 cfu.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.28

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.