

# CERTIFICATE OF ANALYSIS

**PRODUCT** SR0233A  
 SPUTASOL (LIQUID)  
  
**LOT NUMBER** 6267136  
  
**EXPIRY DATE** 2026.05.14  
  
**DATE OF MANUFACTURE** 2025.05.14

**Delivery/Customer information**

Date Printed

2025.06.16

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Colourless liquid	Colourless liquid
pH (25°C)	7.4	7.2 - 7.6

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

**Tested as a diluent**
**Aerobic incubation at 37 ± 2°C for 24 hours**

<i>Pseudomonas aeruginosa</i> ATCC®27853	109	117	107	Straw cols
<i>Branhamella catarrhalis</i> ATCC®25238	164	139	85	Pale straw cols
<i>Staphylococcus aureus</i> ATCC®6538	192	217	113	Straw cols
<i>Haemophilus influenzae</i> ATCC®10211	162	179	110	Pale straw cols
<i>Streptococcus pyogenes</i> ATCC®19615	151	136	90	Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	147	138	94	Grey/green cols, alpha haemolysis
<i>Klebsiella pneumoniae</i> ATCC®700603	125	131	105	Straw cols

**Aerobic incubation at 20-25°C for up to 5 days**

<i>Candida albicans</i> ATCC®10231	36	50	139	Cream, domed cols
------------------------------------	----	----	-----	-------------------

Control Media: Tryptone Soya Agar, Columbia Blood Agar Base enriched with 5% v/v horse blood or Sabouraud Dextrose Agar where appropriate

For *Candida albicans* ATCC®10231, a satisfactory result is represented by recovery greater than or equal to 10% of the Control cfu (0 minutes) from an inoculum of 10-100 colony-forming units (cfu) after holding at 20-25°C for 45 minutes.

For all other strains, a satisfactory result is represented by recovery greater than or equal to 10% of the Control cfu (0 minutes) from an inoculum of 50-200 cfu after holding at 20-25°C for 45 minutes.

Refer to product specification for full details.

**OXOID LIMITED**

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

**CERTIFICATE OF ANALYSIS**

<b>PRODUCT</b>	<b>SR0233A</b>
	<b>SPUTASOL (LIQUID)</b>
<b>LOT NUMBER</b>	<b>6267136</b>

---

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.06.16

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.