

CERTIFICATE OF ANALYSIS

PRODUCT	PB5049A	STAPH/STREP SELEKT.MEDIUM (COLUMBIA CNA)
LOT NUMBER	6270762	
EXPIRY DATE	2025.07.09	
TEST DATE	2025.05.15	
REPORTING DATE	2025.05.19	

General Characteristics	Results	Specification
Colour	Conforms	Traffic red
Appearance	Conforms	Opaque
pH	7.5	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Streptococcus pyogenes ATCC®12344	66	72	1-2mm, grey shiny cols & beta haemolysis
Escherichia coli ATCC®25922	1E+04 - 1E+05	0	Complete inhibition (<= 10 cfu)
Pseudomonas aeruginosa ATCC®27853	1E+04 - 1E+05	0	Complete inhibition (<= 10 cfu)
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, grey cols & alpha haemolysis
Staphylococcus aureus ATCC®25923	<10000	Conforms	Good growth, white shiny colonies

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.