

## CERTIFICATE OF ANALYSIS

<b>PRODUCT</b>	<b>PB5067A</b>	<b>G.VAGINALIS AGAR + HUMAN BLOOD</b>
<b>LOT NUMBER</b>	6281039	
<b>EXPIRY DATE</b>	2025.07.24	
<b>TEST DATE</b>	2025.06.12	
<b>REPORTING DATE</b>	2025.06.16	

General Characteristics	Results	Specification
Colour	Conforms	Signal red
Appearance	Conforms	Opaque
pH	7.3	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Gardnerella vaginalis ATCC®14018	70	56	0.5-1mm, grey cols with haemolysis
Proteus mirabilis ATCC®29906	>100,000	Conforms	No growth
Escherichia coli ATCC®25922	1E+04 - 1E+05	0	Total inhibition
Candida albicans ATCC®10231	1E+04 - 1E+05	0	Total inhibition

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.