

CERTIFICATE OF ANALYSIS

PRODUCT	PB5202E	COLUMBIA AGAR SH.BL./CHOCOLATE AGAR
LOT NUMBER	6283797	
EXPIRY DATE	2025.08.26	
TEST DATE	2025.06.17	
REPORTING DATE	2025.06.20	

General Characteristics Side 1	Results	Specification
Colour	Conforms	Carmine red
Appearance	Conforms	Opaque
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check

Microbiological Performance Side 1	Control c.f.u	Test Result	Specification
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, alpha haemolysis
Streptococcus pyogenes ATCC®12344	<10000	Conforms	Good growth, beta haemolysis

General Characteristics Side 2	Results	Specification
Colour	Conforms	Signal brown
Appearance	Conforms	Opaque
pH	7.3	7.3 -7.7
Packaging / Presentation	Conforms	Label & Print check

Microbiological Performance Side 2	Control c.f.u	Test Result	Specification
Haemophilus influenzae ATCC®10211	<10000	Conforms	Good growth, cream shiny colonies
Neisseria gonorrhoeae ATCC®49226	<10000	Conforms	Good growth, small colonies

3 Day - Contamination Check	Results	Specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.