

CERTIFICATE OF ANALYSIS

PRODUCT	PB5202E	COLUMBIA AGAR SH.BL./CHOCOLATE AGAR
LOT NUMBER	6313379	
EXPIRY DATE	2025.11.12	
TEST DATE	2025.09.03	
REPORTING DATE	2025.09.08	

General Characteristics Side 1	Results		Specification
Colour	Conforms		Carmine red
Appearance	Conforms		Opaque
pH	7.3		7.1 -7.5
Packaging / Presentation	Conforms		Label & Print check
Microbiological Performance Side 1	Control c.f.u	Test Result	Specification
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, alpha haemolysis
Streptococcus pyogenes ATCC®12344	<10000	Conforms	Good growth, beta haemolysis
General Characteristics Side 2	Results		Specification
Colour	Conforms		Signal brown
Appearance	Conforms		Opaque
pH	7.4		7.3 -7.7
Packaging / Presentation	Conforms		Label & Print check
Microbiological Performance Side 2	Control c.f.u	Test Result	Specification
Haemophilus influenzae ATCC®10211	<10000	Conforms	Good growth, cream shiny colonies
Neisseria gonorrhoeae ATCC®49226	<10000	Conforms	Good growth, small colonies
3 Day - Contamination Check	Results		Specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms		Within Limits

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.