

CERTIFICATE OF ANALYSIS

PRODUCT PB5204E SCHAED.SHEEP BL./SCHAEDLER KV H. BLOOD

LOT NUMBER 6268169

EXPIRY DATE 2025.07.17 **TEST DATE** 2025.05.08 **REPORTING DATE** 2025.05.12

General Characteristics Side 1	Results		Specification	
Colour	Conforms		Signal red	
Appearance	Conforms		Opaque	
pH	7.8		7.4 -7.8	
Packaging / Presentation	Conforms		Label & Print check	
Microbiological Performance Side 1	Control c.f.u	Test Result	Specification	
Bacteroides fragilis ATCC®25285	<10000	Conforms	Good growth	
Peptostreptococcus anaerobius ATCC®27337	<10000	Conforms	Good growth	
Clostridium perfringens ATCC®13124	<10000	Conforms	Good growth	
General Characteristics Side 2	Results		Specification	
Colour	Conforms		Vermilion/ Pearl copper	
Appearance	Conforms		Transparent	
pH	7.8		7.4 - 7.8	
Packaging / Presentation	Conforms		Label & Print check	
Microbiological Performance Side 2	Control c.f.u	Test Result	Specification	
Bacteroides fragilis ATCC®25285	<10000	Conforms	Good growth	
Clostridium perfringens ATCC®13124	>100,000	Conforms	No growth	
Escherichia coli ATCC®25922	>100,000	Conforms	No growth	
Peptostreptococcus anaerobius ATCC®27337	>100,000	Conforms	No growth	
3 Day - Contamination Check	Results		Specification	
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms		Within Limits	

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

The quality control methods meet requirements of ISO 11133.





CERTIFICATE OF ANALYSIS

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of $\,$ ISO 11133.

