

G & M Procter Ltd. Certificate of Analysis

PRODUCT **PO1226A**
BRILLIANCE CRE MEDIUM
1X10 PLATTEN

LOT NUMBER 4508550
EXPIRY DATE 2025.07.24
PACKING DATE 0000.00.00
TEST DATE 2025.05.28
REPORTING DATE 2025.06.01

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Blue / green	Blue / green	SOP 178 Appearance and colour
pH (25°C)	7.1	6.8 - 7.2	SOP 53 pH
Fill Weight 19 ± 2 Grams	Conforms	Conforms	SOP 74 Fill volume weight check
Contamination @ 32 ± 2°C for >= 72 hours	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification. For inhibited organisms, the test medium must show no growth from the stated inoculum.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Acinetobacter baumannii NCTC13420	47	49	White/grey colonies	White/grey colonies	SOP 151 Fertility of Specified Target Organism(s)(Agar)
Klebsiella pneumoniae NCTC13443	53	50	Dark blue/purple cols	Dark blue/purple cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball

Ian Snowball
Quality Manager
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

G & M Procter Ltd. Certificate of Analysis

Inhibited Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Escherichia coli ATCC®25922	10,000 - 100,000	No growth	No growth	SOP 155 Inhibition
Klebsiella pneumoniae ATCC®29665	10,000 - 100,000	No growth	No growth	SOP 155 Inhibition
Enterococcus faecalis NCTC12201	10,000 - 100,000	No growth	No growth	SOP 155 Inhibition
Candida albicans ATCC®10231	10,000 - 100,000	No growth	No growth	SOP 155 Inhibition
Sten maltophilia ATCC®13637	10,000 - 100,000	No growth	No growth	SOP 155 Inhibition

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball

Ian Snowball
Quality Manager
G & M Procter Ltd.