

CERTIFICATE OF ANALYSIS

PRODUCT	PO5012A	TRYPTONE SOYA AGAR (TSA)
LOT NUMBER	6291907	
EXPIRY DATE	2026.01.05	
TEST DATE	2025.07.08	
REPORTING DATE	2025.07.14	

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.3	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Incubation @ 30-35°C up to 72h, aerobic			
Escherichia coli ATCC®8739	90	101	2-10 mm, cream colonies
Staphylococcus aureus ATCC®6538	86	87	1-2 mm, cream shiny colonies
Pseudomonas aeruginosa ATCC®9027	55	56	3-8 mm, green-yellow colonies
Bacillus subtilis ATCC®6633	91	99	3-9 mm, cream colonies
Incubation @ 30-35°C up to 120h, aerobic			
Candida albicans ATCC®10231	77	87	2 mm, cream colonies
Aspergillus brasiliensis ATCC®16404	43	42	10-30 mm white mycelium, black spores
Incubation @ 32°C up to 48h, anaerobic			
Clostridium sporogenes ATCC®19404	85	86	1-2 mm, cream colonies
Incubation @ 20-25°C up to 120h, aerobic			
Candida albicans ATCC®10231	77	77	2 mm, cream colonies
Aspergillus brasiliensis ATCC®16404	43	41	10-30 mm white mycelium, black spores

Colony counts shall be equal to or greater than 70% of the control medium (Tryptone Soya Agar or Sabouraud Dextrose Agar)

Tested in accordance with the harmonised methods described in the current European, United States and Japanese pharmacopoeias for the detection of microorganisms in non-sterile products, microbial enumeration tests.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.