

CERTIFICATE OF ANALYSIS

PRODUCT	PO5024C	TSA W/DISINHIBITOR CONTACT
LOT NUMBER	6294479	
EXPIRY DATE	2025.11.06	
TEST DATE	2025.07.18	
REPORTING DATE	2025.07.23	

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.3	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Incubation @ 30-35°C up to 72h, aerobic			
Escherichia coli ATCC®8739	80	53	3-5mm, transparent colonies
Staphylococcus aureus ATCC®6538	68	79	1-3mm, orange shiny colonies
Pseudomonas aeruginosa ATCC®9027	86	60	2-5mm, cream colonies
Bacillus subtilis ATCC®6633	85	58	3-9mm, cream colonies
Streptococcus pyogenes ATCC®19615	38	36	1-2mm, white shiny colonies
Incubation @ 20-25°C up to 72h, aerobic			
Bacillus subtilis ATCC®6633	85	60	3-9mm, cream colonies
Incubation @ 20-25°C up to 120h, aerobic			
Candida albicans ATCC®10231	88	61	2-3mm, cream colonies
Aspergillus brasiliensis ATCC®16404	34	30	10-30mm white mycelium, black spores

Colony counts shall be equal to or greater than 50% of the control medium (Tryptone Soya Agar or Sabouraud Dextrose Agar)

Growth promotions tests are derived from the United States pharmacopoeia for the microbiological control and monitoring of aseptic processing environments.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.