

## CERTIFICATE OF ANALYSIS

**PRODUCT** PO5027A MANNITOL SALT AGAR (CHAPMAN)  
**LOT NUMBER** 6270840  
**EXPIRY DATE** 2025.07.14  
**TEST DATE** 2025.05.12  
**REPORTING DATE** 2025.05.15

General Characteristics	Results	Specification
Colour	Conforms	Antique pink
Appearance	Conforms	Transparent
pH	7.5	7.3 -7.7
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®6538	86	89	1-2mm, yellow shiny colonies
Staphylococcus aureus ATCC®25923	83	88	1-2mm, yellow shiny colonies
Staphylococcus epidermidis ATCC®12228	103	98	1mm, white shiny colonies
Escherichia coli ATCC®25922	>100,000	Conforms	No growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.