

## CERTIFICATE OF ANALYSIS

**PRODUCT** PO5055A MALT EXTRACT AGAR  
**LOT NUMBER** 6270848  
**EXPIRY DATE** 2025.08.29  
**TEST DATE** 2025.05.12  
**REPORTING DATE** 2025.05.15

General Characteristics	Results	Specification
Colour	Conforms	Ocher yellow
Appearance	Conforms	Transparent
pH	5.4	5.2 -5.6
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Candida albicans ATCC®10231	103	97	2-3mm, cream colonies
Aspergillus brasiliensis ATCC®16404	43	39	10-30mm, white mycelium, black spores

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.