

## CERTIFICATE OF ANALYSIS

PRODUCT PO5073A TRYPTONE SOYA AGAR (25ML)

**LOT NUMBER** 6281073

**EXPIRY DATE** 2025.12.16 **TEST DATE** 2025.06.10 **REPORTING DATE** 2025.06.23

<b>General Characteristics</b>	Results		Specification
Colour	Conforms		Ivory
Appearance	Conforms		Transparent
рH	7.3		7.1 -7.5
Packaging / Presentation	Conforms		Label & Print check
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms		Within Limits
Microbiological Performance	Control c.f.u	Test Result	Specification
Incubation @ 30-35°C up to 72h, aerobic			
Escherichia coli ATCC®8739	62	48	2-10 mm, cream colonies
Staphylococcus aureus ATCC®6538	64	61	1-2 mm, cream shiny colonies
Pseudomonas aeruginosa ATCC®9027	76	71	3-8 mm, green-yellow colonies
Bacillus subtilis ATCC®6633	65	62	3-9 mm, cream colonies
Incubation @ 30-35°C up to 120h, aerobic			
Candida albicans ATCC®10231	86	76	2 mm, cream colonies
Aspergillus brasiliensis ATCC®16404	30	28	10-30 mm white mycelium, black spores
Incubation @ 20-25°C up to 120h, aerobic			·
Candida albicans ATCC®10231	86	81	2 mm, cream colonies
Aspergillus brasiliensis ATCC®16404	30	27	10-30 mm white mycelium, black spores

olony counts shall be equal to or greater than 70% of the control medium (Tryptone Soya Agar or Sabouraud Dextrose Agar)

Tested in accordance with the harmonised methods described in the current European, United States and Japanese pharmacopoeias for the detection of microorganisms in non-sterile products, microbial enumeration tests.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

