

CERTIFICATE OF ANALYSIS

PRODUCT PO5076A CETRIMIDE AGAR
LOT NUMBER 6260952
EXPIRY DATE 2025.07.18
TEST DATE 2025.04.11
REPORTING DATE 2025.04.14

General Characteristics	Results	Specification
Colour	Conforms	Oyster white
Appearance	Conforms	Transparent
pH	7.2	6.9 -7.3
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits
Membrane filter 1	Lot 2303723	Product Code GFS13
Membrane filter 2	Lot 2303723	Product Code GFS14
Membrane filter 3	Lot 1406916	Product Code NG145-0045
	Lot 1413234	Product Code NG145-0045
Membrane filter 4	Lot F3HB24241	Product Code NG02054045

Membrane Filtration	Control c.f.u	Test Result	Specification
Pseudomonas aeruginosa ATCC®10145 WDCM 00024	50 - 120	Conforms	2-4 mm, blue-green, fluorescent colonies
Product Code GFS13	91	89	Lot 2303723
Product Code GFS14	91	92	Lot 2303723
Product Code NG145-0045	91	95	Lot 1406916
Product Code NG145-0045	91	85	Lot 1413234
Product Code NG02054045	91	89	Lot F3HB24241
~~~~~			
Escherichia coli ATCC®25922 WDCM 00013	10 ⁴ - 10 ⁵	Conforms	Total inhibition
Product Code GFS13	10 ⁴ - 10 ⁵	Conforms	Lot 2303723
Product Code GFS14	10 ⁴ - 10 ⁵	Conforms	Lot 2303723
Product Code NG145-0045	10 ⁴ - 10 ⁵	Conforms	Lot 1406916
Product Code NG145-0045	10 ⁴ - 10 ⁵	Conforms	Lot 1413234
Product Code NG02054045	10 ⁴ - 10 ⁵	Conforms	Lot F3HB24241
~~~~~			
Enterococcus faecalis ATCC®29212 WDCM 00087	10 ⁴ - 10 ⁵	Conforms	Total inhibition

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

CERTIFICATE OF ANALYSIS

Product Code GFS13	10 ⁴ - 10 ⁵	Conforms	Lot 2303723
Product Code GFS14	10 ⁴ - 10 ⁵	Conforms	Lot 2303723
Product Code NG145-0045	10 ⁴ - 10 ⁵	Conforms	Lot 1406916
Product Code NG145-0045	10 ⁴ - 10 ⁵	Conforms	Lot 1413234
Product Code NG02054045	10 ⁴ - 10 ⁵	Conforms	Lot F3HB24241

Tested in accordance with ISO 11133

Tested in accordance with the methods described in ISO 7704:2023

The formulation of this medium conforms to ISO 16266

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkkS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.