

## CERTIFICATE OF ANALYSIS

**PRODUCT** PO5076A CETRIMIDE AGAR  
**LOT NUMBER** 6278584  
**EXPIRY DATE** 2025.09.12  
**TEST DATE** 2025.06.10  
**REPORTING DATE** 2025.06.12

| General Characteristics                  | Results     | Specification           |
|--|-------------|-------------------------|
| Colour                                   | Conforms    | Oyster white            |
| Appearance                               | Conforms    | Transparent             |
| pH                                       | 7.2         | 6.9 -7.3                |
| Packaging / Presentation                 | Conforms    | Label & Print check     |
| Cont. check @ 20-25 & 30-35°C for >= 72h | Conforms    | Within Limits           |
| Membrane filter 1                        | Lot 2303723 | Product Code GFS13      |
|  | Lot 2404313 | Product Code GFS13      |
| Membrane filter 2                        | Lot 2303723 | Product Code GFS14      |
| Membrane filter 3                        | Lot 1413234 | Product Code NG145-0045 |

| Membrane Filtration                             | Control c.f.u                     | Test Result | Specification                            |
|---|-----------------------------------|-------------|--|
| Pseudomonas aeruginosa ATCC®10145<br>WDCM 00024 | 50 - 120                          | Conforms    | 2-4 mm, blue-green, fluorescent colonies |
| Product Code GFS13                              | 74                                | 60          | Lot 2303723                              |
| Product Code GFS13                              | 74                                | 38          | Lot 2404313                              |
| Product Code GFS14                              | 74                                | 48          | Lot 2303723                              |
| Product Code NG145-0045                         | 74                                | 41          | Lot 1413234                              |
| ~~~~~   |                                   |             |  |
| Escherichia coli ATCC®25922<br>WDCM 00013       | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Total inhibition                         |
| Product Code GFS13                              | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 2303723                              |
| Product Code GFS13                              | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 2404313                              |
| Product Code GFS14                              | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 2303723                              |
| Product Code NG145-0045                         | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 1413234                              |
| ~~~~~   |                                   |             |  |
| Enterococcus faecalis ATCC®29212<br>WDCM 00087  | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Total inhibition                         |
| Product Code GFS13                              | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 2303723                              |
| Product Code GFS13                              | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 2404313                              |
| Product Code GFS14                              | 10 <sup>4</sup> - 10 <sup>5</sup> | Conforms    | Lot 2303723                              |

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

## CERTIFICATE OF ANALYSIS

Product Code NG145-0045

10<sup>4</sup> - 10<sup>5</sup>

Conforms

Lot 1413234

---

Tested in accordance with ISO 11133

Tested in accordance with the methods described in ISO 7704:2023

The formulation of this medium conforms to ISO 16266

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkks according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.