

## CERTIFICATE OF ANALYSIS

**PRODUCT** PO5076A CETRIMIDE AGAR  
**LOT NUMBER** 6278585  
**EXPIRY DATE** 2025.09.16  
**TEST DATE** 2025.06.10  
**REPORTING DATE** 2025.06.13

General Characteristics	Results	Specification
Colour	Conforms	Oyster white
Appearance	Conforms	Transparent
pH	7.2	6.9 -7.3
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits
Membrane filter 1	Lot 2303723	Product Code GFS13
	Lot 2404313	Product Code GFS13
Membrane filter 2	Lot 2303723	Product Code GFS14
Membrane filter 3	Lot 1413234	Product Code NG145-0045

Membrane Filtration	Control c.f.u	Test Result	Specification
Pseudomonas aeruginosa ATCC®10145 WDCM 00024	50 - 120	Conforms	2-4 mm, blue-green, fluorescent colonies
Product Code GFS13	74	50	Lot 2303723
Product Code GFS13	74	39	Lot 2404313
Product Code GFS14	74	52	Lot 2303723
Product Code NG145-0045	74	45	Lot 1413234
~~~~~			
Escherichia coli ATCC®25922 WDCM 00013	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Total inhibition
Product Code GFS13	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 2303723
Product Code GFS13	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 2404313
Product Code GFS14	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 2303723
Product Code NG145-0045	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 1413234
~~~~~			
Enterococcus faecalis ATCC®29212 WDCM 00087	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Total inhibition
Product Code GFS13	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 2303723
Product Code GFS13	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 2404313
Product Code GFS14	10 <sup>4</sup> - 10 <sup>5</sup>	Conforms	Lot 2303723

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

## CERTIFICATE OF ANALYSIS

Product Code NG145-0045

10<sup>4</sup> - 10<sup>5</sup>

Conforms

Lot 1413234

---

Tested in accordance with ISO 11133

Tested in accordance with the methods described in ISO 7704:2023

The formulation of this medium conforms to ISO 16266

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkks according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.