

CERTIFICATE OF ANALYSIS

PRODUCT PO5096A SAB GLUC AGAR +GENTAM+CHLORAMPH
LOT NUMBER 6289637
EXPIRY DATE 2026.01.03
TEST DATE 2025.07.07
REPORTING DATE 2025.07.11

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	5.5	5.4 -5.8
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Candida albicans ATCC®10231	81	65	2-3mm, white colonies
Aspergillus brasiliensis ATCC®16404	1 Col	Conforms	Good growth, white mycelium, black spores
Escherichia coli ATCC®25922	1E+04 - 1E+05	0	Complete inhibition (<= 10 cfu)
Pseudomonas aeruginosa ATCC®27853	1E+04 - 1E+05	0	Complete inhibition (<= 10 cfu)

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.