

CERTIFICATE OF ANALYSIS

PRODUCT	PO5109A	TBX CHROMOGENIC MEDIUM
LOT NUMBER	6278598	
EXPIRY DATE	2025.09.02	
TEST DATE	2025.06.11	
REPORTING DATE	2025.06.16	

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.1	7.0 -7.4
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Escherichia coli ATCC®25922 WDCM 00013	74	48	2-5 mm, turquoise colonies
Escherichia coli NCTC 13216 WDCM 00202	63	47	2-5 mm, colonies with blue centre
Citrobacter freundii ATCC®43864 WDCM 00006	<10000	Conforms	Inhib. growth, white / green-beige cols
Enterococcus faecalis ATCC®29212 WDCM 00087	>100,000	Conforms	No growth

Biological performance tested in accordance with current version of ISO 11133.

Formulation conforms to ISO 16649 (all parts).

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.