

## CERTIFICATE OF ANALYSIS

PRODUCT PO5109A TBX CHROMOGENIC MEDIUM

**LOT NUMBER** 6294511

**EXPIRY DATE** 2025.10.09 **TEST DATE** 2025.07.18 **REPORTING DATE** 2025.07.22

General Characteristics	Results		Specification
Colour	Conforms		Ivory
Appearance	Conforms		Transparent
pH	7.1		7.0 -7.4
Packaging / Presentation	Conforms		Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms		Within Limits
Microbiological Performance	Control c.f.u	Test Result	Specification
Escherichia coli ATCC®25922 WDCM 00013	65	78	2-5 mm, turquoise colonies
Escherichia coli NCTC 13216 WDCM 00202	92	82	2-5 mm, colonies with blue centre
Citrobacter freundiiATCC®43864 WDCM 00006	<10000	Conforms	Inhib. growth, white / green-beige cols
Enterococcus faecalis ATCC®29212 WDCM 00087	>100,000	Conforms	No growth

Biological performance tested in accordance with current version of ISO 11133.

Formulation conforms to ISO 16649 (all parts).

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

