

CERTIFICATE OF ANALYSIS

PRODUCT	PO5132A	PSEUDOMONAS CFC SELECTIVE AGAR
LOT NUMBER	6270875	
EXPIRY DATE	2025.07.23	
TEST DATE	2025.05.14	
REPORTING DATE	2025.05.19	

General Characteristics	Results	Specification
Colour	Conforms	Oyster white
Appearance	Conforms	Transparent
pH	7.1	6.9 -7.3
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Pseudomonas fluorescens ATCC®13525 WDCM 00115	118	117	2-5 mm, irregular yellow/green colonies
Pseudomonas fragi ATCC®4973 WDCM 00116	81	72	3-6 mm, cream colonies
Escherichia coli ATCC®25922 WDCM 00013	1.6E+04	0	Complete inhibition (<= 10 cfu)

Biological performance tested in accordance with current version of ISO 11133.

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.