

CERTIFICATE OF ANALYSIS

| | | |
|-----------------------|----------------|---------------------------------------|
| PRODUCT | PO5132A | PSEUDOMONAS CFC SELECTIVE AGAR |
| LOT NUMBER | 6281088 | |
| EXPIRY DATE | 2025.08.19 | |
| TEST DATE | 2025.06.11 | |
| REPORTING DATE | 2025.06.16 | |

| General Characteristics | Results | Specification |
|--|----------|---------------------------|
| Colour | Conforms | Oyster white |
| Appearance | Conforms | Transparent |
| pH | 7.1 | 6.9 -7.3 |
| Packaging / Presentation | Conforms | Conforms to specification |
| Cont. check @ 20-25 & 30-35°C for >= 72h | Conforms | Within Limits |

| Microbiological Performance | Control c.f.u | Test Result | Specification |
|--|---------------|-------------|---|
| Pseudomonas fluorescens ATCC®13525 WDCM 00115 | 58 | 64 | 2-5 mm, irregular yellow/green colonies |
| Pseudomonas fragi ATCC®4973 WDCM 00116 | 62 | 46 | 3-6 mm, cream colonies |
| Escherichia coli ATCC®25922 WDCM 00013 | 1.8E+04 | 0 | Complete inhibition (<= 10 cfu) |

Biological performance tested in accordance with current version of ISO 11133.

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.