

**G & M Procter Ltd.  
Certificate of Analysis**

**PRODUCT** PO5133A  
**BACILLUS CEREUS MEDIUM (MYP)**  
**1X10 PLATTEN**

**LOT NUMBER** 4531125  
**EXPIRY DATE** 2026.03.03  
**PACKING DATE** 2025.11.25  
**TEST DATE** 2025.11.26  
**REPORTING DATE** 2025.11.28

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Orange	Orange	SOP 178 Appearance and colour
pH (25°C)	7.4	7.0 - 7.4	SOP 53 pH
Fill Weight 19 ± 2 Grams	Conforms	Conforms	Fill volume weight check
Contamination @ 32 ± 2°C for >= 72 hours	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

**MICROBIOLOGICAL PERFORMANCE**

Tested in accordance with ISO11133:2014. For target organisms the control media must achieve 50-120 cfu. The test medium must achieve between 50% & 120% of the control medium and show the colonial appearance stated in the specification. For non target organisms the test medium must show the colonial appearance or level of inhibition as stated in the specification from the stated inoculums.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Bacillus cereus ATCC®11778	94	74	Bright pink cols with zones	Bright pink cols with zones	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



Ian Snowball  
Quality Manager  
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,  
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

## G & M Procter Ltd. Certificate of Analysis

Inhibited Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Escherichia coli ATCC®25922	100000	Complete inhibition (<= 10 cfu)	Complete inhibition (<= 10 cfu)	SOP 155 Inhibition

Target Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Bacillus subtilis ATCC®6633	10,000 - 100,000	Yellow/orange cols, no zones	Yellow/orange cols, no zones	SOP 199 Colonial appearance (Plates Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



*Ian Snowball*

Ian Snowball  
Quality Manager  
G & M Procter Ltd.