

CERTIFICATE OF ANALYSIS

PRODUCT	PO5159A	BRILLIANCE (TM) UTI CLARITY (TM)
LOT NUMBER	6278470	
EXPIRY DATE	2025.08.25	
TEST DATE	2025.06.03	
REPORTING DATE	2025.06.06	

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.0	6.8 -7.2
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Escherichia coli ATCC®25922	98	84	2-4mm, rose shiny cols, Indol +
Staphylococcus aureus ATCC®25923	108	92	1-2mm, white colonies
Enterococcus faecalis ATCC®29212	<10000	Conforms	Good growth, turquoise shiny colonies
Proteus mirabilis ATCC®29906	<10000	Conforms	Good growth, cream cols brown halo, Ind.-
Klebsiella oxytoca ATCC®13182	<10000	Conforms	Good growth, blue shiny colonies

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.