

CERTIFICATE OF ANALYSIS

PRODUCT	PO5219E	CAMPYL. KARMALI/ CAMPYL. KARMALI
LOT NUMBER	6273828	
EXPIRY DATE	2025.08.21	
TEST DATE	2025.05.16	
REPORTING DATE	2025.05.19	

General Characteristics Side 1	Results	Specification
Colour	Conforms	Jet black
Appearance	Conforms	Opaque
pH	7.4	7.2 -7.6
Packaging / Presentation	Conforms	Conforms to specification

Microbiological Performance Side 1	Control c.f.u	Test Result	Specification
Campylobact. jejuni ssp.jejuni ATCC®33291	<10000	Conforms	Good growth, grey flat shiny colonies
Escherichia coli ATCC®25922	>100,000	Conforms	No growth
Staphylococcus epidermidis ATCC®12228	>100,000	Conforms	No growth

General Characteristics Side 2	Results	Specification
Colour	Conforms	Jet black
Appearance	Conforms	Opaque
pH	7.5	7.2 -7.6
Packaging / Presentation	Conforms	Conforms to specification

Microbiological Performance Side 2	Control c.f.u	Test Result	Specification
Campylobact. jejuni ssp.jejuni ATCC®33291	<10000	Conforms	Good growth, grey flat shiny colonies
Escherichia coli ATCC®25922	>100,000	Conforms	No growth
Staphylococcus epidermidis ATCC®12228	>100,000	Conforms	No growth

3 Day - Contamination Check	Results	Specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

CERTIFICATE OF ANALYSIS

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkkS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.