

CERTIFICATE OF ANALYSIS

PRODUCT	PO5428J	CHROMOGENER COLIFORMEN AGAR(ISO)(55MM)
LOT NUMBER	6286611	
EXPIRY DATE	2025.08.21	
TEST DATE	2025.06.26	
REPORTING DATE	2025.06.30	

General Characteristics	Results	Specification
Colour	Conforms	Light ivory
Appearance	Conforms	Transparent
pH	6.9	6.6 -7.0
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits
Membrane filter 1	Lot 2303723	Product Code GFS13
	Lot 2404313	Product Code GFS13
Membrane filter 2	Lot 2303723	Product Code GFS14
Membrane filter 3	Lot 1413234	Product Code NG145-0045

Membrane Filtration	Control c.f.u	Test Result	Specification
Escherichia coli ATCC®8739 WDCM 00012	50 - 120	Conforms	0,5 - 2 mm, dark blue to violet colonies
Product Code GFS13	81	67	Lot 2303723
Product Code GFS13	81	66	Lot 2404313
Product Code GFS14	81	59	Lot 2303723
Product Code NG145-0045	81	66	Lot 1413234
~~~~~			
Enterobacter aerogenes ATCC®13048 WDCM 00175	50 - 120	Conforms	0,5 - 2 mm, pink to red colonies
Product Code GFS13	94	72	Lot 2303723
Product Code GFS13	94	75	Lot 2404313
Product Code GFS14	94	78	Lot 2303723
Product Code NG145-0045	94	69	Lot 1413234
~~~~~			
Pseudomonas aeruginosa ATCC®10145 WDCM 00024	10 ³ - 10 ⁴	Conforms	Good growth, colourless colonies
Product Code GFS13	10 ³ - 10 ⁴	Conforms	Lot 2303723
Product Code GFS13	10 ³ - 10 ⁴	Conforms	Lot 2404313
Product Code GFS14	10 ³ - 10 ⁴	Conforms	Lot 2303723

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.

CERTIFICATE OF ANALYSIS

Product Code NG145-0045	10 ³ - 10 ⁴	Conforms	Lot 1413234
~~~~~			
Enterococcus faecalis ATCC®19433 WDCM 00009	10 ⁴ - 10 ⁵	Conforms	Total inhibition
Product Code GFS13	10 ⁴ - 10 ⁵	Conforms	Lot 2303723
Product Code GFS13	10 ⁴ - 10 ⁵	Conforms	Lot 2404313
Product Code GFS14	10 ⁴ - 10 ⁵	Conforms	Lot 2303723
Product Code NG145-0045	10 ⁴ - 10 ⁵	Conforms	Lot 1413234

Tested in accordance with ISO 11133 and ISO 9308-1

Formulation compliant with ISO 9308-1

Tested in accordance with the methods described in ISO 7704:2023

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.