

## Certificate of Analysis

**Product Name:** McFarland Turb Standard 2.0  
**Product Code:** R20412  
**Lot Number:** 6226140  
**Expiration Date:** 22.06.2026  
**Date of Manufacture:** 06.05.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W77BCL08J	McFarland Standard 2.0 Filled Bulk	6226169	22.06.2026

McFarland Equivalence Turbidity Standards are designed to be visual standards for determining approximate microorganism concentration.

A 2.0 McFarland Standard is approximately equivalent to 600 million cfus/ml. Variation in organisms size and phase of log growth will affect organism density.



Fay Curd  
Quality Assurance Manager