

Certificate of Analysis

Product Name: McFarland Turb Standard 5.0
Product Code: R20415
Lot Number: 6261131
Expiration Date: 21.09.2026
Date of Manufacture: 04.08.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W77BCL11M	McFarland Standard 5.0 Filled Bulk	6261152	21.09.2026

McFarland Equivalence Turbidity Standards are designed to be visual standards for determining approximate microorganism concentration.

A 5.0 McFarland Standard is approximately equivalent to 1.5 billion cfus/ml. Variation in organisms size and phase of log growth will affect organism density.



Fay Curd
Quality Assurance Manager