

Certificate of Quality

Product Name: Coagulase Plasma 25ml/VL
Lot Number: 269365


Product Number: R21052
Expiration Date: 2026-11-05
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|----------------------------|-------|----------------------------------|
| Staphylococcus aureus | 25923 | (+) Tight clot at 4 and 24 Hours |
| Staphylococcus aureus | 33862 | (+) Tight clot at 4 and 24 Hours |
| Staphylococcus aureus | 6538 | (+) Tight clot at 4 and 24 Hours |
| Staphylococcus aureus | 29213 | (+) Tight clot at 4 and 24 Hours |
| Staphylococcus epidermidis | 12228 | (-) No clot at 4 and 24 Hours |

Physical Characteristics:
Appearance: White To Light Beige, Intact Cake
pH @ 20-25°C: N/A

Microbial Load Evaluation:
Product is evaluated for performance specifications only.

Signed


Senior Quality Assurance Engineer