

Certificate of Quality

Product Name: BactiDrop Oxidase PK/50
Lot Number: 256754


Product Number: R21540
Expiration Date: 2027-01-09
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Neisseria sicca	9913	(+) Purple, Oxidase Positive
Pasteurella multocida	43137	(+) Purple, Oxidase Positive
Pseudomonas aeruginosa	27853	(+) Purple, Oxidase Positive
Escherichia coli	25922	(-) No Color Change, Oxidase Negative

Physical Characteristics:
Appearance: Colorless to Light Pink / Beige, Clear Liquid
pH @ 20-25°C: N/A

Microbial Load Evaluation:
Product is evaluated for performance specifications only.

Signed

Senior Quality Assurance Engineer