

Certificate of Quality

Product Name: BactiDrop Spot Indole PK/50
Lot Number: 270088


Product Number: R21550
Expiration Date: 2027-03-11
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides ovatus	8483	(+) Blue, Indole Positive
Escherichia coli	25922	(+) Blue, Indole Positive
Prevotella melaninogenica	25845	(-) Pink, Indole Negative
Proteus mirabilis	12453	(-) Pink, Indole Negative

Physical Characteristics:
Appearance: Yellow, Clear Liquid
pH @ 20-25°C: N/A

Microbial Load Evaluation:
Product is evaluated for performance specifications only.

Signed

Senior Quality Assurance Engineer