

## Certificate of Analysis

**Product Name:** ProSpecT C. diff Toxin A/B MP KT/96TST  
**Product Code:** R244596  
**Lot Number:** 6236607  
**Expiration Date:** 05.03.2026  
**Date of Manufacture:** 03.07.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
ELY1104025	STOP SOLUTION 12ML	6175815	31.07.2029
ELY1106048	GEN WASH 120ML	6175756	30.11.2026
W75PCS01S	Colour Substrate TMB-W 25ml	6128470	31.12.2026
W75PCJ14U	C Diff A/B Conjugate 25ml	6217177	05.03.2026
W75PPL14B	C Diff A/B Plate 96 Test	6140525	30.04.2026
W75PPC13U	C Diff A/B Pos Control	6217178	06.06.2027
W75PNC05F	C Diff A/B Neg Control	6217179	04.06.2027
W75PSB04D	C Diff A/B Diluent 120ml	6217180	03.05.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Specification	Result
Negative Control	Visually Negative	Pass
	Mean OD < = to 0.080	Pass
Positive Control	Visually Positive	Pass
	Mean OD > = to 1.000	Pass
Positive Samples	Visually Positive	Pass
	Mean OD > = to 0.130	Pass
Negative Samples	Visually Negative	Pass
	Mean OD < 0.080	Pass



Fay Curd  
Quality Assurance Manager