

Certificate of Analysis


Product Name: ProSpecT C. diff Toxin A/B MP KT/96TST
Product Code: R244596
Lot Number: 6287421
Expiration Date: 14.07.2026
Date of Manufacture: 02.09.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75PCJ14U	C Diff A/B Conjugate 25ml	6266378	14.07.2026
ELY1104025	STOP SOLUTION 12ML	3774595	30.09.2028
ELY1106048	GEN WASH 120ML	6175756	30.11.2026
W75PCS01S	Colour Substrate TMB-W 25ml	6128470	31.12.2026
W75PPL14B	C Diff A/B Plate 96 Test	6301713	21.05.2027
W75PPC13U	C Diff A/B Pos Control	6266379	14.10.2027
W75PNC05F	C Diff A/B Neg Control	6266380	14.10.2027
W75PSB04D	C Diff A/B Diluent 120ml	6266381	14.09.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Specification	Result
Negative Control	Visually Negative Mean OD < = to 0.080	Pass Pass
Positive Control	Visually Positive Mean OD > = to 1.000	Pass Pass
Positive Samples	Visually Positive Mean OD > = to 0.130	Pass Pass
Negative Samples	Visually Negative Mean OD < 0.080	Pass Pass



Fay Curd
Quality Assurance Manager