

Certificate of Analysis

Product Name: ProSpecT Giardia MP KT/96TST
Product Code: R2458096
Lot Number: 6248932
Expiration Date: 31.03.2026
Date of Manufacture: 03.06.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75PCJ03S	Giardia Conjugate 25ml	6225541	31.03.2026
W75PCS01S	Colour Substrate TMB-W 25ml	6128470	31.12.2026
W75PNC01C	Generic Negative Control	6212681	13.02.2027
W75PPC03Y	Giardia Pos Control	6162874	02.04.2027
W75PPL03D	Giardia Plate 96 Test	3804997	31.03.2027
W75PSB01U	General Sample Diluent 120ml	6195484	13.02.2027
ELY1104025	STOP SOLUTION 12ML	6175815	31.07.2029
ELY1106048	GEN WASH 120ML	6212686	31.01.2027

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Specification	Result
Negative Control	Visually Negative Mean OD < = to 0.100	Pass Pass
Positive Control	Visually Positive Mean OD > = to 0.300	Pass Pass (blanked)
Positive Samples	Visually Positive Mean OD > = to 0.05	Pass Pass (blanked)
Negative Samples	Visually Negative Mean OD < 0.05	Pass Pass (blanked)



Fay Curd
Quality Assurance Manager