

Certificate of Analysis

Product Name: Haemophilus influenzae Type c 2ml/VL
Product Code: R30166201
Lot Number: 6228790
Expiration Date: 13.05.2028
Date of Manufacture: 22.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75BZM22X	H. Influenzae Type C Filled Bulk	6222068	13.05.2028

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Identity & Counterimmuno Electrophoresis	A positive reaction must be seen with each suspension and its homologous antiserum	Conforms
	The Serum under test must give a positive reaction with its homologous suspensions	Conforms
	There must not be any reaction between wells containing saline and suspension.	Conforms
Identity Slide Agglutination	The Homologous controls must show agglutination equal to or greater than +	Conforms
	The Serum under test must show agglutination equal to or greater than + with homologous cultures	Conforms
	The Saline control should not give agglutination although a trace is acceptable	Conforms



Fay Curd
 Quality Assurance Manager