

Certificate of Analysis

Product Name: Neisseria meningitidis Gp A 2ml/VL
Product Code: R30166801
Lot Number: 6223337
Expiration Date: 31.03.2028
Date of Manufacture: 29.04.2025

*Date format is DD.MM.YYYY

| Item | Description | Lot/Serial | Expire |
|-----------|------------------------------------|------------|------------|
| W74AZM37V | N.Meningitidis Group A Filled Bulk | 6205056 | 31.03.2028 |

Test results carried out according to the Quality Assurance Product Release and Test Specification

| Test | Anticipated Result | Actual Result |
|--|--|---------------|
| Abbreviated Identity Slide Agglutination | Homologous control must give agglutination equal to or greater than + | Conforms |
| | Serum under test must show agglutination equal to or greater than + with the homologous culture | Conforms |
| | Serum under test must not show agglutination with the heterologous culture, trace reaction is acceptable | Conforms |
| | Saline control should not give agglutination, although a trace is acceptable | Conforms |
| Slide agglutination test using positive and negative cultures | Homologous control must give agglutination equal to or greater than + | Conforms |
| | Serum under test must show agglutination equal to or greater than + with the homologous (positive) culture | Conforms |
| | Serum under test must not show agglutination with the heterologous (negative) culture | Conforms |
| | Saline control should not give agglutination, although a trace is acceptable | Conforms |



Fay Curd
 Quality Assurance Manager