

Certificate of Analysis

Product Name: Salmonella O Poly (Gp A - S) 2ml/VL
Product Code: R30858201
Lot Number: 6233855
Expiration Date: 15.04.2027
Date of Manufacture: 04.07.2025

*Date format is DD.MM.YYYY

| Item | Description | Lot/Serial | Expire |
|-----------|-------------------------------------|------------|------------|
| W75AZC02A | Salmonella Poly-O (A-S) Filled Bulk | 6196414 | 15.04.2027 |

Test results carried out according to the Quality Assurance Product Release and Test Specification:

| Test | Anticipated result | Actual Result |
|---|---|---------------|
| Abbreviated Identity Slide Agglutination | Homologous control wells must show agglutination equal to or greater than + | Conforms |
| | Negative control well must not show agglutination (trace is acceptable) | Conforms |
| | Serum under test must give agglutination with all cultures equal to or greater than + | Conforms |



Fay Curd
 Quality Assurance Manager