

Certificate of Analysis

Product Name: Staphaurex KT/400 TST
Product Code: R30859902
Lot Number: 6234318
Expiration Date: 11.09.2026
Date of Manufacture: 01.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74TZL30M	Staphaurex Latex Filled Bulk	6199917	11.09.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Smoothness	Test latex must not show agglutination although a trace granularity is acceptable	Conforms
	Test Latex should spread easily to cover the circle	Conforms
Efficacy/ Specificity	Test Latex must show obvious agglutination with the appropriate positive Staphylococcus cultures	Conforms
	Test Latex must not show agglutination with the appropriate negative Staphylococcus cultures (trace granularity is acceptable)	Conforms



Fay Curd
Quality Assurance Manager